



**SIGNET 8860 METER**  
**WATER QUALITY METER SETTINGS**

INSTALLED IN RO SERIAL NO. \_\_\_\_\_ MODEL NO. \_\_\_\_\_

SHIPPED TO \_\_\_\_\_

**CH. 1 CELL STANDARD:** .1  
**CH. 1 SET:** CONDUCTIVITY  
**CH. 1 UNITS:** PPM

**RELAY 1 HYS:** 0.5000  
**RELAY 1 DELAY:** 0.0 SECS  
**RELAY 2 MODE:** HIGH

**CH. 1 TDS:** 2.0000 uS/PPM

**RELAY 2 SOURCE:** CH. 2 COND.

**CH. 2 CELL STANDARD:** .1

**RELAY 2 SETPOINT:** 0.0000 - 30.0000 PPM  
DAVITA  
0.000 - 50.0000 PPM  
ALL OTHERS

**CH. 2 SET:** CONDUCTIVITY

**RELAY 2 HYS:** 0.5000 PPM

**CH. 2 UNITS:** PPM

**RELAY 2 DELAY:** 0.0 SECS.

**CH. 2 TDS:** 2.0000 Us/ PPM

**RELAY 3 MODE:** OFF

**FUNCTION:** REJECT C1? C2

**RELAY 4 MODE:** OFF

**LOOP 1 SOURCE:** CH.1 COND.

**LAST CAL.:** ENTER CURRENT DATE  
\_\_\_\_/\_\_\_\_/\_\_\_\_

**LOOP 1 RNG:** PPM 0.0000 – 30.0000 PPM FOR DAVITA  
0.0000 – 50.0000 PPM ALL OTHERS

**CALIBRATED BY:** \_\_\_\_\_

**LOOP 2 SOURCE:** CH. 2 COND.

**LOOP 2 RNG:** PPM 0.0000 – 30.0000 PPM FOR DAVITA  
0.0000 – 50.0000 PPM ALL OTHERS

**COMMENTS:** \_\_\_\_\_

**LOOP 3 SOURCE:** FUNCTION – DO NOT CAL.

**LOOP 3 RNG:** % - DO NOT CAL.

**RELAY 1 MODE:** HIGH

**RELAY 1 SOURCE:** CH. 2 COND.

**RELAY 1 SET POINT:** 0.0000 – 30.0000 PPM FOR DAVITA  
0.0000 – 50.0000 PPM ALL OTHERS